

# VELVET GRIP

Disposable NITRILE GLOVES



curaden  
better health for you

## CERTIFICATIONS

cura **GRIP**<sup>+</sup>  
A SWISS BRAND [www.curagrip.com](http://www.curagrip.com)

<b>INDEX:</b>	<b>01</b>	<b>MAIN PAGE</b> .....	1
	<b>02</b>	<b>INDEX</b> .....	2
	<b>03</b>	<b>CERTIFICATES</b> .....	3
	<b>04</b>	<b>ISO CERTIFICATES</b>	3
		• <b>ISO9001:2005</b> .....	3
		• <b>ISO 13485: 2016</b> .....	4
		<b>EN 455</b>	
		• <b>EN 455:1-4</b> .....	5
		<b>EN 374</b>	
		• <b>EN 374-1-2, EN ISO 374-4-5, EN 21420</b> .....	10
		• <b>CE 0197</b> .....	17
		<b>EN 1186</b> .....	18
		<b>FDA</b> .....	19
		<b>DoC</b> .....	20
	<b>05</b>	<b>PACKAGING</b> .....	22
	<b>06</b>	<b>LAST PAGE</b> .....	23



## MANUFACTURING ACCREDITATIONS

### ISO 9001: 2015



## MANUFACTURING ACCREDITATIONS

### ISO 13485:2016



### Certificate

**Quality Management System**  
EN ISO 13485:2016

Registration No.:  
Organization:

Scope: Design and Development, Manufacture and Distribution of Patient Examination Gloves

The Certification Body of TÜV Rheinland LGA Products GmbH certifies that the organization has established and applies a quality management system for medical devices. Proof has been furnished that the requirements specified in the abovementioned standard are fulfilled. The quality management system is subject to yearly surveillance.

Report No.:  
Effective date: 2021-04-15  
Expiry date: 2024-04-14  
Issue date: 2021-04-13






### Certificate

**Quality Management System**  
EN ISO 13485:2016

Registration No.:  
Organization:

The scope of certification also covers the following:

No.	Facility	Scope
/01		Distribution of Patient Examination Gloves
/02		Design and Development, Manufacture of Patient Examination Gloves

Report No.:  
Effective date: 2021-04-15  
Expiry date: 2024-04-14  
Issue date: 2021-04-13






### Certificate

**Quality Management System**  
EN ISO 13485:2016

Registration No.:  
Organization:

The scope of certification also covers the following:

/03	Design and Development, Manufacture of Patient Examination Gloves
/04	Design and Development, Manufacture of Patient Examination Gloves

Report No.:  
Effective date: 2021-04-15  
Expiry date: 2024-04-14  
Issue date: 2021-04-13






**Business Stream Products**  
Certification Department

TÜV Rheinland LGA Products GmbH • 91108 Kitzingen

Contact:  
Tel: +49 91 939 9229  
Mail: service@lga.com  
Date: April 13, 2021

Application for:  
Certificate No.:  
Requirement: EN ISO 13485:2016

Dear Madam or Sir,  
Enclosed please find the new certificate No. I replacing the previous certificate.  
With effective date of the new certificate, the previous certificate becomes invalid.

Best regards,  
  
Jing Zhang  
Certification body

TÜV Rheinland LGA Products GmbH  
Am Gassen 10a  
91108 Kitzingen  
Germany  
Telephone: +49 91 939 9229  
Fax: +49 91 939 9220  
www.lga.com  
E-Mail: service@lga.com

Report No.:  
Effective date: 2021-04-15  
Expiry date: 2024-04-14  
Issue date: 2021-04-13






## INTERNATIONAL STANDARDS

### EN 455: 1-4

**Test Report No.**  
dated 20 Oct 2020



**NOTE:** This report is issued subject to the testing and certification regulations of the TUV SUD Group and the General Terms and Conditions of Business of TUV SUD which may be found at www.tuv.com or in German at the same or on request to the client.

**SUBJECT:**

**TESTED FOR:**

**TEST DATE:**  
22 Jul 2020 to 14 Aug 2020, 20 Oct 2020


**DESCRIPTION OF SAMPLES:**

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Nitrile Disposable Exam Gloves	Blue	S	S	400	

Lot size as specified by client: 150,001 to 500,000 pieces

**METHOD OF TEST:**


- EN 455-1:2020 Medical gloves for single use  
Part 1: Requirements and testing for freedom from holes
- EN 455-2:2015 Medical gloves for single use  
Part 2: Requirements and testing for physical properties
- EN 455-3:2015 Medical gloves for single use  
Part 3: Requirements and testing for biological evaluation
  - Clause 4.4 Powder-free gloves
  - Clause 4.6 Labelling




TUV SUD PAS  
 Laboratory:  
 TUV SUD PAS/PLA  
 HL Strasse 119-121  
 D-82031 Rosenheim  
 Germany 18231

Phone: +49 89 303 323  
 Fax: +49 89 303 324  
 E-mail: info@tuv.com  
 www.tuv.com (in German)  
 CA No. 18032013

Registered Office:  
 TUV SUD Area Centre AG  
 1 Strasse 119-121  
 D-82031 Rosenheim  
 Germany 18231  
**TUV**



**Test Report No.**  
dated 20 Oct 2020



**RESULTS:**

Sample: Nitrile Disposable Exam Gloves, Lot No. , Blue, Size S

**Table 1: Results for EN 455-1:2020**


Clause	Tests	Requirements	No. of non-conformities allowed (pieces)	Number tested (pieces)	Actual no. of non-conformities found (pieces)	Inferred results
4	Freedom from holes	Shall not leak	10	315	0	Passed

**Table 2: Results for EN 455-2:2015 Clauses 4-5**


Clause	Tests	Requirements (Median)	Number tested (pieces)	Results (Median)	Inferred results
4	Dimensions a) Length (mm)	≥ 240	13	246	Passed
	b) Width (mm)	For Size S: 80 ± 10	13	83	Passed
5	Strength at break (N)	For nitrile examination gloves: ≥ 6.0	13	7.6	Passed
	b) Force at break after challenge testing (N) 7 days at 70±2°C	For nitrile examination gloves: ≥ 6.0	13	7.4	Passed

**Table 3: Results for EN 455-2:2015 Clause 7**

Clause	Tests	Requirements	Results	Inferred results
7	Labelling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1:2012, and EN 1041:2008(A1):2013. Date of manufacture is defined on the packaging date.	Observed	Passed



**Test Report No.**  
dated 20 Oct 2020



**RESULTS (cont'd):**


Sample: Nitrile Disposable Exam Gloves, Lot No. , Blue, Size S

**Table 4: Results for EN 455-3:2015 Clause 4.4**

Clause	Tests	Requirements	Results / Remarks	Inferred results
4.4	Powder-free gloves	For powder-free gloves, The total quantity of powder residues shall not exceed 2 mg per glove.	0.52 mg per glove	Passed
5.2				

**Table 5: Results for EN 455-3:2015 Clause 4.6**

Clause	Tests	Requirements	Results
4.6	Labelling	In addition to the labelling specified in EN 1041:2008+A1:2013 and the relevant symbols given in EN ISO 15223-1:2012, the following requirements apply:	
		a) medical gloves containing natural rubber latex shall be labelled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1:2012 symbol for latex.	NA
		The labelling shall include the following or equivalent warning statement together with the symbol: (Product) contains natural rubber latex which may cause allergic reactions, including anaphylactic responses.	NA
		b) the labelling shall include a prominent indication of whether the glove is powdered or powder-free.	Comply
		c) sterile powdered gloves shall be labelled with the following or equivalent: CAUTION. Surface powder shall be removed specifically prior to undertaking operative procedures in order to minimize the risk of adverse tissue reactions.	NA
		d) for any medical glove containing natural rubber latex the product labelling shall not include: <ul style="list-style-type: none"> <li>- any term suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein;</li> <li>- any unqualified indication of the presence of allergen.</li> </ul>	NA
e) if the manufacturer labels the gloves with the protein content, the process limit, measured as specified in 8.3, shall be given.	NA		
		Inferred results	Passed



**Test Report No.**  
dated 20 Oct 2020



**REMARKS:**

- Labelling requirements are assessed based on submitted packaging artwork by client on 20 Oct 2020.
- NA: Not applicable for the submitted sample.

Yao Jiah Kuang  
 Associate Engineer

  
 Le Qile Yi  
 Engineer  
 Medical Health Services (NMM)

**APPENDIX:**

**Photo 1:** Nitrile Disposable Exam Gloves, Lot No.



**Photo 2:** Packaging Artwork for Nitrile Disposable Exam Gloves, Lot No.




## INTERNATIONAL STANDARDS

### EN 455: 1-4

**Test Report No.**  
dated 20 Oct 2020




**Please note that this Report is issued under the following terms:**

- The report applies to the sample of the specific production/shipment given at the time of its realization. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD approves, recommends or endorses the manufacturer, supplier or user of such production/shipment, or that TÜV SÜD PSE in any way "guarantees" the safe performance of the production/shipment. Unless otherwise stated in this report, no tests were conducted to determine long-term effects of using the specific production/shipment.
- The samples mentioned in this report were submitted/produced/manufactured by the Client. TÜV SÜD PSE therefore assumes no responsibility for the accuracy of information on the listed name, model number, origin of manufacture, compliance or any information supplied.
- Nothing in this report shall be interpreted to mean that TÜV SÜD PSE has verified or ascertained any endorsement or marks from any other body/authority or bodies that may be found on the sample.
- This report shall not be reproduced wholly or in part, and no reference shall be made by the Client to TÜV SÜD PSE or to the report or results furnished by TÜV SÜD PSE in any other form or in any other context.
- Unless otherwise stated, the tests were carried out in TÜV SÜD PSE File No. 1 Standard Risk Class Singapore (11021).
- The tests carried out by TÜV SÜD PSE and the report are subject to TÜV SÜD PSE's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SÜD Group.

Effective 01 September 2020




**Test Report No.**  
dated 20 Oct 2020



**Subject:**

**TESTED FOR:**

**TEST DATE:**  
22 Jul 2020 to 14 Aug 2020, 00 Oct 2020


**DESCRIPTION OF SAMPLES:**

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Nitrile Disposable Exam Gloves	Blue		M	400	


Lot size as specified by Client: 150,000 to 500,000 pieces

**METHOD OF TEST:**

- EN 455-1:2010 Medical gloves for single use  
Part 1: Requirements and testing for freedom from holes
- EN 455-2:2010 Medical gloves for single use  
Part 2: Requirements and testing for physical properties
- EN 455-3:2015 Medical glove for single use  
Part 3: Requirements and testing for biological evaluation  
- Clause 4.4 Powder-free gloves  
- Clause 4.6 Labelling

**Test Report No.**  
dated 20 Oct 2020



**RESULTS:**

Sample: Nitrile Disposable Exam Gloves, Lot No.

**Table 1: Results for EN 455-1:2010**


Clause	Tests	Requirements (objectives)	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4	Freedom from holes	Shall not leak	10	315	3	Passed

**Table 2: Results for EN 455-2:2010, Clause 4.5**


Clause	Tests	Requirements (objectives)	Number tested (pieces)	Results (pieces)	Inferred results
4	a) Length (mm)	≥ 240	13	245	Passed
	b) Width (mm)	For Size M: 98 ± 10	13	93	Passed
5	a) Force at break (N)	For nitrile examination gloves: ≥ 6.0	13	6.9	Passed
	b) Force of break after challenge testing (N) 7 days at 170±2°C	For nitrile examination gloves: ≥ 6.0	13	7.1	Passed

**Table 3: Results for EN 455-2:2010, Clause 7**

Clause	Tests	Requirements	Results	Inferred results
7	Labelling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1:2012 and EN ISO 1041:2008+A1:2013. Date of manufacture is defined as the packaging date.	Observed	Passed



**Test Report No.**  
dated 20 Oct 2020



**RESULTS (cont'd):**


Sample: Nitrile Disposable Exam Gloves, Lot No.

**Table 4: Results for EN 455-3:2015, Clause 4.4**

Clause	Tests	Requirements	Results / Remarks	Inferred results
4.4	Powder-free gloves	For powder-free gloves: The total quantity of powder residues shall not exceed 2 mg per glove.	1.08 mg per glove	Passed

**Table 5: Results for EN 455-3:2015, Clause 4.6**

Clause	Tests	Requirements	Results
4.6	Labelling	In addition to the labelling specified in EN ISO 1041:2008+A1:2013 and the relevant symbols given in EN ISO 15223-1:2012, the following requirements apply: a) medical gloves containing natural rubber latex shall be labelled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1:2012 symbol for latex. The labelling shall include the following or equivalent warning statement together with the symbol: "Product contains natural rubber latex which may cause allergic reactions, including anaphylactic responses"; b) the labelling shall include a prominent indication of whether the glove is powder-free or powder-free; c) sterile powdered gloves shall be labelled with the following or equivalent: "CAUTION: Surface powder shall be removed aseptically prior to undertaking operative procedures in order to minimize the risk of adverse tissue reactions"; d) for any medical glove containing natural rubber latex the product labelling shall not include: - any term suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein; - any unqualified indication of the presence of allergens; e) if the manufacturer labels the gloves with the protein content, the protein limit, measured as specified in 6.3, shall be given.	NA
		Comply	
		NA	
		Inferred results	Passed



## INTERNATIONAL STANDARDS

### EN 455:1-4

**Test Report No.**  
dated 20 Oct 2020



**REMARKS:**

- Labeling requirements are assessed based on submitted packaging artwork by client on 20 Oct 2020.
- NA: Not applicable for the submitted sample.

*Yoo Poh Kwang*  
Associate Engineer

*Wong Chai Yi*  
Engineer  
Medical Health Services (NAM)

**APPENDIX:**



Photo 1: Nitrile Disposable Exam Gloves, Lot No.



Photo 2: Packaging Artwork for Nitrile Disposable Exam Gloves, Lot No.



**Test Report No.**  
dated 20 Oct 2020




**Please note that this Report is issued under the following terms:**

- This report applies to the sample of the specific production/shipment given at the time of its designation. The results are not valid to indicate or imply that they are applicable to other items. In addition, such results shall not be used to indicate or imply that TÜV SÜD PSE approves, recommends or endorses the manufacturer, supplier or user of such production/shipment, or that TÜV SÜD PSE in any way guarantees the safe performance of the production/shipment. Please reference stated in this report, no tests were conducted to determine long term effects using the specific production/shipment.
- The samples mentioned in this report shall remain the property of the Client. TÜV SÜD PSE therefore assumes no responsibility for the accuracy of information on the label, name, model number, origin of manufacture, consignment or any information supplied.
- Nothing in this report shall be interpreted to mean that TÜV SÜD PSE has verified or guaranteed any endorsement or marks from any other testing authority or bodies that may be found on that sample.
- This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSE or to the report contents furnished by TÜV SÜD PSE in any advertisements or sales promotion.
- Unless otherwise stated, the tests were carried out in TÜV SÜD PSE (Pte. Ltd., No.1 Selegie Road, Singapore 11827).
- The tests carried out by TÜV SÜD PSE and this report are subject to TÜV SÜD PSE's General Terms and Conditions of Business and the Testing/Certification Regulations of the TÜV SÜD Group.

Effective 1 September 2020




**Test Report No.**  
dated 20 Oct 2020



**Note:** This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSE (Pte. Ltd.). In addition, this report is governed by the terms stated within this report.

**SUBJECT:**  
All values, test results.

**TESTED FOR:**

**TEST DATE:**  
22 Jul 2020 to 14 Aug 2020, 20 Oct 2020



**DESCRIPTION OF SAMPLES:**

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Nitrile Disposable Exam Gloves	Blue		L	399	

Lot size as specified by client: 150,000 to 500,000 pieces

**METHOD OF TEST:**

- EN 455-1:2010 Medical gloves for single use  
Part 1: Requirements and testing for leaeson from holes
- EN 455-2:2015 Medical gloves for single use  
Part 2: Requirements and testing for physical properties
- EN 455-3:2015 Medical glove for single use  
Part 3: Requirements and testing for biological evaluation  
- Clause 4.4 Powder-free gloves  
- Clause 4.6 Labeling


**Laboratory:**  
TUV SÜD PSE  
No. Selegie Road  
Singapore 11827

**Phone:** +65 6886 1122  
**Fax:** +65 6711 6001  
**E-mail:** enquiries@tuv.com  
**Web:** www.tuv.com  
**Co. Reg.:** 196230258

**Registered Office:**  
TUV SÜD Asia Pacific Pte. Ltd.  
Chongqing Rd. 20, #02-01  
Singapore 11827

Page 1 of 3

**Test Report No.**  
dated 20 Oct 2020



**RESULTS:**

Sample: Nitrile Disposable Exam Gloves, Lot No.

**Table 1: Results for EN 455-1:2010**



Cause	Tests	Requirements	No. of non-complies allowed (pieces)	Number tested (pieces)	Actual no. of non-complies found (pieces)	Inferred results
4	Freedom from holes	Shall not leak	10	315	3	Passed

**Table 2: Results for EN 455-2:2015 Clauses 4-5**

Cause	Tests	Requirements (Median)	Number tested (pieces)	Results (Median)	Inferred results
4	a) Length (mm)	≥ 240	13	243	Passed
	b) Width (mm)	For Size L: 175 ± 10	13	104	Passed
5	a) Strength at Force at break (N)	For nitrile examination gloves: ≥ 8.0	13	7.0	Passed
	b) Force at break after challenges testing (N)	For nitrile examination gloves: ≥ 6.6	13	8.1	Passed

**Table 3: Results for EN 455-2:2015 Clause 7**

Cause	Tests	Requirements	Results	Inferred results
7	Labeling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1:2012 and EN 1041:2008+A1:2013. Date of manufacture is defined as the packaging date.	Observed	Passed





Page 2 of 3

## INTERNATIONAL STANDARDS

### EN 455:1-4

Test Report No. dated 20 Oct 2020



**RESULTS (cont'd):**


Sample: Nitrile Disposable Exam Gloves, Lot No.

Table 4: Results for EN 455-2:2015 Clause 4.4

Clause	Tests	Requirements	Results / Remarks	Inferred results
4.4	Powder-free gloves	For powder-free gloves, The total quantity of powder residues shall not exceed 2 mg per glove.	0.96 mg per glove	Passed

Table 5: Results for EN 455-3:2015 Clause 4.6

Clause	Tests	Requirements	Results
4.6	Labelling	In addition to the labelling specified in EN 1041:2008+A1:2013 and the relevant symbols given in EN ISO 15223-1:2012, the following requirements apply:	NA
		a) medical gloves containing natural rubber latex shall be labelled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1:2012 symbol for latex.	NA
		The labelling shall include the following or equivalent warning statement together with the symbol: "Product contains natural rubber latex which may cause allergic reactions, including anaphylactic responses."	NA
		b) the labelling shall include a prominent indication of whether the glove is powdered or powder-free.	Comply
		c) sterile powdered gloves shall be labelled with the following or equivalent: "CAUTION: Surface powder shall be removed aseptically prior to undertaking operative procedures in order to minimize the risk of adverse tissue reactions."	NA
d) for any medical glove containing nitrile rubber latex the product labelling shall not exclude: <ul style="list-style-type: none"> <li>- any term suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein;</li> <li>- any unqualified indication of the presence of allergen;</li> </ul>	NA		
e) if the manufacturer labels the gloves with the protein content, the protein limit, measured as specified in 5.3 shall not apply.	NA		
Inferred results			Passed



Test Report No. dated 20 Oct 2020



**REMARKS:**

- Labelling requirements are assessed based on submitted packaging artwork by client on 20 Oct 2020.
- NA: Not applicable for the submitted sample.

Yeo Poh Kwang Associate Engineer

Yeo DN Yi Engineer Medical Health Services (NAAL)

**APPENDIX:**



Photo 1: Nitrile Disposable Exam Gloves, Lot No.



Photo 2: Packaging Artwork for Nitrile Disposable Exam Gloves, Lot No.



Test Report No. dated 20 Oct 2020




**Please note that this Report is issued under the following terms:**

- This report applies to the sample of the specific production/shipment given at the time of its testing/calibration. The results are not valid to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply the TÜV SUD PSD approval, assessment or approval by manufacturer, supplier or user of test production/shipment, or the TÜV SUD PSD in any way whatsoever. The later performance of the production/shipment, supplier or user of test production/shipment, or the TÜV SUD PSD in any way whatsoever is not covered by this report. No tests were conducted to determine long-term effects of using the specific production/shipment.
- The samples mentioned in this report belong to the client and are not to be used for any other purpose. The client is responsible for the accuracy of information on the test name, material number, origin of manufacture, composition or any information required.
- Nothing in this report shall be interpreted to mean that TÜV SUD PSD has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on the sample.
- This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SUD PSD or the report or results furnished by TÜV SUD PSD in any advertisements or other publications.
- Unless otherwise stated, the tests were carried out in TÜV SUD PSD (the UK, has) 15 Chester Drive, Chislehurst, Essex, SSM 7 6LJ.
- The tests carried out by TÜV SUD PSD and this report are subject to TÜV SUD PSD's General Terms and Conditions of Business and the Testing and Calibration Regulations of the TÜV SUD Group.

Effective 31 September 2021




Test Report No. dated 20 Oct 2020



**Note:** This report is subject to the Testing and Calibration Regulations of the TÜV SUD Group and the General Terms and Conditions of Business of TÜV SUD PSD (the UK). In addition, this report is governed by the terms set out within this report.

**SUBJECT:**

**TESTED FOR:**

**TEST DATE:**  
22 Jul 2020 to 14 Aug 2020, 20 Oct 2020

**DESCRIPTION OF SAMPLES:**

SN	Product Description	Colour	Lot No.	Size	Sample received (quantity)	Manufacturer
1	Nitrile Disposable Exam Gloves	Blue		XL	401	

Lot size as specified by client: 150,001 to 500,000 pieces

**METHOD OF TEST:**


- EN 455-1:2020 Medical gloves for single use  
Part 1: Requirements and testing for freedom from holes
- EN 455-2:2015 Medical gloves for single use  
Part 2: Requirements and testing for physical properties
- EN 455-3:2015 Medical glove for single use  
Part 3: Requirements and testing for biological evaluation  
Clause 4.4 Powder-free gloves  
Clause 4.6 Labelling

**TUV SUD PSD**

Labortary: 15 Chester Drive, Chislehurst, Essex, SSM 7 6LJ

Phone: +44 (0) 20 8500 7000  
Fax: +44 (0) 20 8500 7000  
Email: enquiry@tuvatuk.com  
www.tuvatuk.com

Reg and Prod Order: 15 Chester Drive, Chislehurst, Essex, SSM 7 6LJ  
15 Chester Drive, Chislehurst, Essex, SSM 7 6LJ






## INTERNATIONAL STANDARDS

### EN 455:1-4

**Test Report No.**  
dated 20 Oct 2020



**RESULTS:**  
Sample: Nitrile Disposable Exam Gloves, Lot No.

**Table 1: Results for EN 455:1:2012**


Clause	Tests	Requirements	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4.5	Freedom from holes	Shall not leak	10	315	2	Passed

**Table 2: Results for EN 455:2:2015 Clauses 4-5**


Clause	Tests	Requirements (Median)	Number tested (pieces)	Results (Median)	Inferred results
4	a) Length (mm)	≥ 240	13	240	Passed
	b) Width (mm)	For Size XL: ≥ 110	13	114	Passed
	Strength at Force at break (N)	For nitrile examination gloves: ≥ 6.0	13	6.8	Passed
5	b) Force of break after challenge testing (N) 7 days at 70±2°C	For nitrile examination gloves: ≥ 4.0	13	7.0	Passed

**Table 3: Results for EN 455:3:2015 Clause 7**

Clause	Tests	Requirements	Results	Inferred results
7	Labeling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1:2012 and EN ISO 1041:2008+A1:2013. Date of manufacture is defined on the packaging data.	Observed	Passed



**Test Report No.**  
dated 20 Oct 2020




**RESULTS (cont'd):**  
Sample: Nitrile Disposable Exam Gloves, Lot No.      Size XL

**Table 4: Results for EN 455:3:2015 Clauses 4.4**

Clause	Tests	Requirements	Results / Remarks	Inferred results
4.4.5.2	Powder-free gloves	For powder-free gloves: The total quantity of powder residues shall not exceed 2 mg per glove.	0.04 mg per glove	Passed

**Table 5: Results for EN 455:3:2015 Clause 4.6**

Clause	Tests	Requirements	Results
4.6	Labeling	In addition to the labeling specified in EN 1041:2008+A1:2013 and the relevant symbols given in EN ISO 15223-1:2012, the following requirements apply: a) medical gloves containing natural rubber latex shall be labeled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1:2012 symbol for latex. The labeling shall include the following or equivalent warning statement together with the symbol: (Product) contains natural rubber latex which may cause allergic reactions, including anaphylactic responses. b) the labeling shall include a prominent indication of whether the glove is powdered or powder-free. c) sterile powdered gloves shall be labeled with the following or equivalent: CAUTION: Surface powder shall be removed meticulously prior to undertaking operative procedures in order to minimize the risk of adhesive tissue reactions. d) for any medical glove containing natural rubber latex the product labeling shall not include: - any term suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein; - any unqualified indication of the presence of allergens. e) if the manufacturer labels the gloves with the protein content, the protein limit, measured as specified in 5.3 shall be given.	Comply
		Inferred results	Passed



**Test Report No.**  
dated 20 Oct 2020



**REMARKS:**

- Labeling requirements are assessed based on submitted packaging artwork by client on 20 Oct 2020.
- NA: Not applicable for the submitted sample.

Yee Poh Kwang  
Associate Engineer

Yee Poh Kwang  
Engineer  
Medical Health Services (MAH)

**APPENDIX:**



Photo 1: Nitrile Disposable Exam Gloves, Lot No.



Photo 2: Packaging Artwork for Nitrile Disposable Exam Gloves, Lot No.



**Test Report No.**  
dated 20 Oct 2020



**Please note that this Report is issued under the following terms:**

- The report applies to the sample of the specific product/procedure given at the time of its testing/operation. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results shall not be used to establish or imply that TÜV SUD PSE approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SUD PSE in any way "guarantees" the safe performance of the product/equipment. Users are advised to refer to the manufacturer's instructions to determine the proper use of any specific product/equipment.
- The services mentioned in this report were conducted by the manufacturer of the Client. TÜV SUD PSE neither assumes nor is responsible for the accuracy of information or the brand name, model number, origin of manufacturer, design or any information supplied.
- Nothing in this report shall be interpreted to mean that TÜV SUD PSE has verified or ascertained any endorsement or marks from any other leading authority or bodies that may be found on that sample.
- The report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SUD PSE or to the report or results furnished by TÜV SUD PSE in any advertisement or sales promotion.
- Unless otherwise stated, the tests were carried out in TÜV SUD PSE, Pte Ltd, No. 1 Science Park Drive, Singapore 118221.
- The tests carried out by TÜV SUD PSE and this report are subject to TÜV SUD PSE's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SUD Group.


Effective September 2020




## INTERNATIONAL STANDARDS

### EN 374-1-2, EN 374-4-5, EN 21420

Prüfbericht - Produkte Test Report - Products		TÜVRheinland®	
Prüfbericht-Nr.: Test report no.:	Module C2 Annual Check	Auftrags-Nr.: Order no.:	2022-01-28
Kunden-Referenz-Nr.: Client reference no.:	Module C2 Annual Check	Auftragsdatum: Order date:	2022-01-28
Auftraggeber: Client:			
Prüfgegenstand: Test item:	Schutzhandschuhe / Protective gloves		
Bezeichnung / Typ-Nr.: Identification / Type no.:	Nitrilhandschuhe Nitrile gloves		
Auftrags-Inhalt: Order content:	Produktüberwachung entsprechend Modul C2 der Verordnung (EU) 2016/425 product control according to Regulation (EU) 2016/425 of module C2		
Prüfgrundlage: Test specification:	EN ISO 374-1:2016 + A1:2018 Schutzhandschuhe gegen gefährliche Chemikalien und Mikroorganismen Protective gloves against dangerous chemicals and micro-organisms		
Wareneingangdatum: Date of sample receipt:	2022-04-22		
Prüfmuster-Nr.: Test sample no.:			
Prüfzeitraum: Testing period:	2022-05-25 - 2022-06-13		
Ort der Prüfung: Place of testing:	Prüfstelle für Textilien und PSA Köln		
Prüflaboratorium: Testing laboratory:	TÜV Rheinland LGA Products GmbH		
Prüfresultat: Test result:	Pass		
geprüft von: tested by:		genehmigt von: authorized by:	
Datum: Date:	2022-06-13	Ausstellungsdatum: Issue date:	2022-06-13
Stellung / Position: Sachverständigen/Expert	Sachverständigen/Expert	Stellung / Position: Sachverständigen/Expert	Sachverständigen/Expert
Sonstiges / Other:	Prüfungen ausgewählter Parameter entsprechend o.g. Prüfgrundlage im Rahmen der Produktüberwachung gemäß Modul C2 der Verordnung (EU) 2016/425 Tests of selected parameters according to the above mentioned test basis within the scope of product monitoring in accordance with module C2 of Regulation (EU) 2016/425		
Zustand des Prüfgegenstandes bei Anlieferung: Condition of the test item at delivery:	Prüfmuster vollständig und unbeschädigt Test item complete and undamaged		
<p><small>*Typen: - nitril - entspricht o.g. Prüfgrundlage / - nitril - entspricht o.g. Prüfgrundlage / - nitril - entspricht o.g. Prüfgrundlage / - nitril - entspricht o.g. Prüfgrundlage</small></p> <p><small>**Typen: - nitril - entspricht o.g. Prüfgrundlage / - nitril - entspricht o.g. Prüfgrundlage / - nitril - entspricht o.g. Prüfgrundlage / - nitril - entspricht o.g. Prüfgrundlage</small></p> <p>Dieser Prüfbericht bezieht sich nur auf das o.g. Prüfmuster und darf ohne Genehmigung der Prüfstelle nicht auszugeben verwendet werden. Dieser Bericht berechtigt nicht zur Verwendung eines Prüfmusters. This test report only relates to the above mentioned test sample. Without permission of the test center, the test report is permitted to be duplicated in extracts. This test report does not enable in any way test results to be used for other test samples.</p> <p>TÜV Rheinland LGA Products GmbH Tilsiterstr. 20 - 40401 Nürnberg Mail: service@lga.tuv.com Web: www.lga.com</p>			



Prüfbericht - Produkte Test Report - Products		TÜVRheinland®	
Prüfbericht-Nr.: Test report no.:		Seite 2 von 28 Page 2 of 28	
<b>Anmerkungen</b> Remarks			
<p>1. Alle eingesetzten Prüfmittel waren zum angegebenen Prüfzeitraum gemäß eines festgelegten Kalibrierungsprogramms unseres Prüflabors kalibriert. Sie entsprechen den in den Prüfprogrammen hinterlegten Anforderungen. Die Rückverfolgbarkeit der eingesetzten Prüfmittel ist durch die Einhaltung der Regelungen unseres Managementsystems gegeben. Detailed information regarding test conditions, equipment and measurement uncertainty is available in the test laboratory and could be provided on request.</p> <p>The equipment used during the specified testing period was calibrated according to our test laboratory calibration program. The equipment fulfills the requirements included in the relevant standards. The traceability of the test equipment used is ensured by compliance with the regulations of our management system. Detailed information regarding test conditions, equipment and measurement uncertainty is available in the test laboratory and could be provided on request.</p> <p>2. Wie vertraglich vereinbart, wurde dieses Dokument nur digital unterzeichnet. Der TÜV Rheinland hat nicht überprüft, welche rechtlichen oder sonstigen zusätzlichen Anforderungen für dieses Dokument gelten. Diese Überprüfung liegt in der Verantwortung des Benutzers dieses Dokuments. Auf Verlangen des Kunden kann der TÜV Rheinland die Gültigkeit der digitalen Signatur durch ein gesondertes Dokument bestätigen. Diese Anfrage ist an unseren Vertrieb zu richten. Eine Umweltgebühr für einen solchen zusätzlichen Service wird erhoben. As contractually agreed, this document has been signed digitally only. TÜV Rheinland has not verified and is unable to verify which legal or other pertaining requirements are applicable for this document. Such verification is within the responsibility of the user of this document. Upon request by the client, TÜV Rheinland can confirm the validity of the digital signature by a separate document, such request should be addressed to our Sales department. An environmental fee for such additional service will be charged.</p> <p>3. Prüfklausel mit der Note "1" wurden an qualifizierte Unterlieferanten vergeben und sind unter der jeweiligen Prüfklausel des Berichts beschrieben. Abweichungen von Prüfspezifikationen) oder Kundenanforderungen sind in der jeweiligen Prüfklausel im Bericht aufgeführt. Test clauses with remark of "1" are subcontracted to qualified subcontractors and described under the respective test clause in the report. Deviations of testing specification(s) or customer requirements are listed in specific test clause in the report.</p> <p>4. Die Entscheidungsregel für Konformitätsfeststellungen in diesem Prüfbericht basiert auf der "Null-Quotenregel" und der "Einfachen Akzeptanz" gemäß ILAC G8:2019 und IEC Guide 1:2021, es sei denn, in der auf Seite 1 dieses Berichts genannten Norm ist etwas anderes festgelegt oder vom Kunden gewünscht. Dies bedeutet, dass die Messunsicherheit nicht berücksichtigt wird und daher auch nicht im Prüfbericht angegeben wird. The decision rule for statements of conformity in this test report is based on the "Zero Quota Band Rule" and "Simple Acceptance" in accordance with ILAC G8:2019 and IEC Guide 1:2021, unless otherwise specified in the applied standard mentioned on Page 1 of this report or requested by the customer. This means that measurement uncertainty is not taken in account and hence also not declared in the test report.</p> <p>5. Vorhersehbare Verwendung wurde betrachtet. Zurzeit liegen für das die Produkt's weder Schutzlassverfahren an, noch ist ein erhöhtes Unfallrisiko bekannt. Foreseeable use was considered. Currently neither a safeguard clause procedure has been implemented nor an increase in accidents known for this product is.</p>			



Prüfbericht - Produkte Test Report - Products		TÜVRheinland®	
Prüfbericht-Nr.: Test report no.:		Seite 3 von 28 Page 3 of 28	
<b>Anmerkungen</b> Remarks			
<b>Überwachungshistorie / Surveillance history</b>			
EU Baumusterzertifikat Nr.: EU Type Certificate No.:	I / vom / dated 2021-03-18		
Jahr / Year	geprüfte Abschneite / tested Clauses		
2019	EN 420, 5.1.2, 5.2; EN ISO 374-2, 7.1, 7.2, 7.3; EN ISO 374-4, 5.3, 5.3.4, 5.4, EN ISO 374-5, 5.1, 5.2		
2020	EN 420, 4.1, 4.2, 4.3, 5; EN ISO 374-2, 7.1, 7.2, 7.3; EN ISO 374-4, 5.3, 5.3.4, 5.4, 5.5, EN 420, 6, 7		
2021	EN ISO 374-2 Teil 7, 5.3; EN ISO 374-4, 5; EN ISO 374-5, 5.6, 7; EN ISO 21420 Teil 4, Teil 7		
2022	EN 374-2 Teil 7, EN ISO 374-4 5.4, 5.5, EN ISO 374-5 5.1, 5.2, 5.4, Teil 6 und 7, EN 21420 Teil 4, 5 und 7		



Prüfbericht - Produkte Test Report - Products		TÜVRheinland®	
Prüfbericht-Nr.: Test report no.:		Seite 4 von 28 Page 4 of 28	
<b>Produktbeschreibung</b> Product description			
1. Produktdetails Product details	5-Finger-Handschuh 5 finger gloves		
2. Artikel / Modell Article / Model	Nitrilhandschuhe Nitrile gloves		
3. Größe / Länge Size / Length	S (6,5), M (7,5), L (8,5), XL (9)		
4. Leistungsstufen Performance levels	Chemikale K: NaOH 40%	Permeation: Klasse/level 6	Degradation: -4,3 %
5. Verwendete Materialien Used materials	Nitril Nitrile Farbe/colour blau Materialdicke 0,07 mm waf/ thickness 0,07 mm		
6. Pflegekennzeichnung/ care instruction	n/a		
7. Mitgelieferte Dokumente / Prüfberichte Further applicable documents / test reports	*1 Prüfbericht Unsicherheitskoeffizient / Test report inaccuracy Bericht-Nr. / report no. of 2022-06-09		
8. Sonstiges Other	Test sample(s), as well sample information, description, product details and intended usage was provided by customer.		
9. Prüfmusterbereitstellung Test sample obtaining	☐ Sending by customer ☑ Sampling by TÜV Rheinland Group Datum Musterziehung/sample picking date: 2022-03-31		
Verpackung/packaging			



## INTERNATIONAL STANDARDS

### EN 374-1-2, EN 374-4-5, EN 21420




Prüfbericht - Produkte  
Test Report - Products

Prüfbericht-Nr.: \_\_\_\_\_ Seite 6 von 28  
Test report no.: \_\_\_\_\_ Page 6 of 28

Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
Der Originaltext wird nur auszugsweise wiedergegeben. Details sind dem Original-Dokument zu entnehmen. The original text is reproduced only in part. For details, be referred to the original document.			
1	Anwendungsbereich Scope		
2	Normative Verweisungen Normative references		
3	Begriffe Terms and definition		
4	Probennahme Sampling		
5	Leistungsanforderung Performance requirements		
5.1	Allgemeine Anforderungen General requirements	Schutzhandschuhe gegen gefährliche Chemikalien müssen die Anforderungen in EN ISO 21420:2020, Abschnitt 4, Absatz 11 und Abschnitt 7, erfüllen. Protective gloves against dangerous chemicals shall comply with the requirements given in EN ISO 21420:2020, Clause 4, Clause 5 and Clause 7.	
5.2	Penetration Penetration	Schutzhandschuhe dürfen bei der Prüfung nach EN 374-2:2019, 7.2 und 7.3, nicht undicht werden. Protective gloves shall not leak when tested according to EN 374-2:2019, 7.2 and 7.3.	
EN 374-2	Teil 2: Bestimmung des Widerstandes gegen Penetration Part 2: Determination of resistance to penetration		
4.3	Bemerkungen Remarks	Das Luft-Leck-Verfahren ist nicht für alle Handschuhe geeignet. Beispielsweise können die Teile einiger Handschuhe zu stark aufblasen sein, während andere Teile derselben Handschuhe nur teilweise aufblasen sein können. Wenn sich die Luft-Leck-Prüfung als ungeeignet erweist, dann wird nur die Prüfung auf Penetration von Wasser durchgeführt. Bei beiden Verfahren werden keine Undichtigkeiten berücksichtigt, die bis zu 40 mm vom Rand des Flüssigkeitsdurchdringlichen Bereichs des Handschuhes entfernt liegen. The air leak procedure is not suitable for all gloves. For example parts of some gloves may be overinflated while other parts of the same gloves can only be partially inflated. If the air leak test proves unsuitable, then only the water penetration test is carried out. For both methods dispersed leaks within the area of 40 mm from the edge of the liquid proof area.	
7	Durchführung Procedure		
7.1	Allgemeines General	Der Handschuh wird vorsichtig der Hülle, Schachtel oder seiner Verpackung entnommen. Identifizierungscode, Nummer des Loses, Größe und Marke der Proben werden aufgeschrieben. Eine Schprüfung auf Risse, Schlitze und Löcher wird durchgeführt. Sind diese vorhanden, ist anzugeben, dass die Handschuhe die Prüfung nicht bestanden haben.	

TPP: EN ISO 374-1:2016 + A1:2018 (08/20) - Y1.0







Prüfbericht - Produkte  
Test Report - Products

Prüfbericht-Nr.: \_\_\_\_\_ Seite 6 von 28  
Test report no.: \_\_\_\_\_ Page 6 of 28

Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
Carefully remove the glove from the wrapper, box or its packaging, record the identity code, lot number, size and brand of samples. Visually examine for tears, rips and holes. If these are present, the gloves shall be reported as failing batch.			
7.2	Luft-Leck-Prüfung Air leak test		no tears, rips and holes are present
4.1	Ein Handschuh wird in Wasser getaucht und sein Innenleben mit Luft aufgeblasen. Eine Undichtheit (Leck) wird als Strom aus Luftblasen sichtbar, der sich an der Oberfläche des Handschuhes bildet. A glove is immersed in water, and its interior is pressurized with air. A leak is detected by a stream of air bubbles from the surface of the glove.	Größe/ size Luft-Leck-Prüfung/ Air leakage S keine/no Leakage M keine/no Leakage L keine/no Leakage XL keine/no Leakage	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
Tab. 1	Nenngröße der Handschuhe (e) nach Angaben des Herstellers (e) Nominal glove thickness (e) in mm As provided by the manufacturer.	Luftdruck (X) Air pressure (X) kPa	Verwendeter Luftdruck / air pressure user: 0.5 kPa
	e = 0.3	0.3	
	0.3 < e < 0.5	2.0	
	0.5 < e < 1.0	5.0	
	e > 1.0	6.0	
7.3	Wasser-Leck-Prüfung Water leak test		
4.2	Ein Handschuh wird mit Wasser gefüllt. Eine Undichtheit wird durch das Auftreten von Wassertröpfchen an der Außenseite des Handschuhes festgestellt. A glove is filled with water. A leak is detected by the appearance of water droplets on the outside of the glove.	Größe/ size Wasser-Leck-Prüfung / Water leakage S keine/no Leakage M keine/no Leakage L keine/no Leakage XL keine/no Leakage	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
5.3	Degradation Degradation		P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
5.4	Permeation Permeation		
5.4.1	Allgemeines General	Für den Handschuh, der länger als 400 mm ist, und bei dem die Handinnenfläche und die Sohle unterschiedliche Leistungsstufen erreichen, muss für jede Chemikalie die geringere Leistungsstufe in der Kennzeichnung angegeben werden. Alle Ergebnisse sollen in der Benutzeranleitung angegeben sein. Für jede Kombination von Schutzhandschuh-Prüfchemikalien ist nach Tabelle 1 zu klassifizieren, wobei die EN 16023-1:2015, 8.5.1.1 oder 8.5.1.3, angegebene Ergebnisse für die normalisierte Darstellung anzuwenden sind. For the glove longer than 400 mm, where the palm and cuff achieve different performance levels, the low performance level shall be claimed in the marking for each chemical. All the results of tests reported to the user (instruction).	

TPP: EN ISO 374-1:2016 + A1:2018 (08/20) - Y1.0






Prüfbericht - Produkte  
Test Report - Products


Prüfbericht-Nr.: \_\_\_\_\_ Seite 7 von 28  
Test report no.: \_\_\_\_\_ Page 7 of 28

Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
Each combination of protective glove/test chemical shall be classified according to Table 1, using the results as given in EN 16023-1:2015, 8.5.1.1 or 8.5.1.3 for the normalised breakthrough time.			
5.4.2	Type A: Die Permeationsleistung muss mindestens Stufe 2 gegen wenigstens sechs Prüfchemikalien entsprechen, die in Tabelle 2 gelistet sind. Type A: The permeation performance shall be at least level 2 against a minimum of six test chemicals listed in Table 2.		
5.4.3	Type B: Die Permeationsleistung muss mindestens Stufe 2 gegen wenigstens drei Prüfchemikalien entsprechen, die in Tabelle 2 gelistet sind. Type B: The permeation performance shall be at least level 2 against minimum of three test chemicals listed in Table 2.		
5.4.3	Type C: Die Permeationsleistung muss mindestens Stufe 1 gegen wenigstens sechs Prüfchemikalien entsprechen, die in Tabelle 2 gelistet ist. Type C: The permeation performance shall be at least level 1 against minimum of six test chemical listed in Table 2.		
Tab. 2	Kennbuchstabe Code Letter	Prüfchemikalie Chemical	CAS-RN CAS Number
	A	Methanol / Methanol	67-56-1
	B	Aceton / Acetone	62-64-1
	C	Acetonitril / Acetonitrile	75-05-8
	D	Dichlormethan / Dichloromethane	75-69-2
	E	Kohlensäure / carbon dioxide	78-09-3
	F	Toluol / Toluene	108-88-3
	G	Dibutylzinn / Dibutyltin	109-59-7
	H	Tetrahydrofuran / Tetrahydrofuran	109-99-9
	J	Ethylacetat / Ethyl acetate	141-78-6
	L	Diäthyl- <i>n</i> -Hexane	142-82-6
	K	Natriumhydroxid 40 % / Sodium hydroxide 40 %	1312-73-2
	L	Schwefelsäure 96 % / Sulphuric acid 96 %	7664-93-9
	M	Salzsaure 35 % / Nitric acid 35 %	7697-37-2
	N	Essigsäure 99 % / Acetic acid 99 %	64-19-7
	O	Ammoniumhydroxid 20 % / Ammonium hydroxide 20 %	1336-21-6
	P	Wasserstoffperoxid 30 % / Hydrogen peroxide 30 %	7724-84-1
	S	Fluorsäure 40 % / Hydrofluoric acid 40 %	7664-39-3
	T	Formaldehyd 37 % / Formaldehyde 37 %	50-00-0
Leistungsstufen gegen Permeation Permeation performance level			
Tab. 1	Gemessene Durchbruchzeit / Measured breakthrough time (min)	Schutzindex / Protection performance level	Prüfchemikalie / Chemical
	> 10	Klasse / class 1	Natriumhydroxid 40 % / Sodium hydroxide 40 %
	> 30	Klasse / class 2	Sodium hydroxide 40 %
	> 60	Klasse / class 3	
	> 120	Klasse / class 4	
	> 240	Klasse / class 5	
	> 480	Klasse / class 6	

Die Prüfchemikalien (muss / müssen) aus der Liste der Prüfchemikalien in Tabelle 2 genommen werden.

TPP: EN ISO 374-1:2016 + A1:2018 (08/20) - Y1.0






Prüfbericht - Produkte  
Test Report - Products

Prüfbericht-Nr.: \_\_\_\_\_ Seite 8 von 28  
Test report no.: \_\_\_\_\_ Page 8 of 28

Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result			
Abhängig von der Anwendung der Handschuhe könnten andere Prüfchemikalien verwendet werden. The test chemicals shall be taken from the list of test chemicals in Table 2. Other test chemicals could be used depending on the application of the gloves.						
5.5	Anforderungen an Handschuh-Typen A, B und C Requirements for different protection types of gloves					
Tab. 3						
		5.1	5.2	5.4.2	5.4.3	5.4.4
	Type A / Type A	X	X	X	X	X
	Type B / Type B	X	X	X	X	X
	Type C / Type C	X	X	X	X	X
	X = erforderlich / required					
EN ISO 374-5	Teil 5: Terminologie und Leistungsanforderungen für Risiken durch Mikroorganismen Part 5: Terminology and performance requirements for micro-organisms risks					
5	Leistungsanforderung Performance requirement					
5.1	Allgemeine Anforderungen General requirement	Schutzhandschuhe gegen Mikroorganismen sollen der EN ISO 21420:2020, Absatz 4, Abs. 5 und Abs. 7 entsprechen. Protective gloves against micro-organisms shall comply with the requirements given in EN ISO 21420:2020, Clause 4, Clause 5 and Clause 7.	gegeben / given			P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
5.2	Penetration Penetration	Schutzhandschuhe gegen Viren, Bakterien und Pilze dürfen bei der Prüfung nach EN 374-2:2019, 7.2 und 7.3 nicht undicht werden. Protective gloves against virus, bacteria and fungi shall not leak when tested according to EN 374-2:2019, 7.2 and 7.3.	gegeben, siehe Abschnitt 7.2 und 7.3			P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
5.3	Schutz vor Viren Protection against viruses	Schutzhandschuhe gegen Viren sind nach ISO 16604 Verfahren B zu testen und dürfen im Testzylinder keinen nachweisbaren Transfer (< 1 PFU/ml) des Phi-X174-Bakteriophagen aufweisen. Protective gloves against virus shall be tested according to ISO 16604 Procedure B and shall exhibit no detectable transfer (< 1 PFU/ml) of the Phi-X174 bacteriophage in the assay tube.	nicht anwendbar, da nicht spezifiziert			not applicable, because not marked

TPP: EN ISO 374-1:2016 + A1:2018 (08/20) - Y1.0



## INTERNATIONAL STANDARDS

### EN 374-1-2, EN 374-4-5, EN 21420

**Prüfbericht - Produkte**  
Test Report - Products

**TÜVRheinland**<sup>®</sup>

Prüfbericht-Nr.: Seite 9 von 28  
Test report no.: Page 9 of 28

Abzatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result												
5.4	Anforderungen an verschiedene Schutzarten von Handschuhen Requirements for different protection types of gloves Die Anforderungen sind in der Tabelle 1 aufgeführt. The requirements are mentioned in the Table 1.														
Tab. 1	<table border="1"> <thead> <tr> <th></th> <th>5.1</th> <th>5.2</th> <th>5.3</th> </tr> </thead> <tbody> <tr> <td>Handschuh gegen Bakterien und Pilze (Glove protection against bacteria and fungi)</td> <td>X</td> <td>X</td> <td></td> </tr> <tr> <td>Handschuh gegen Viren, Bakterien und Pilze (Glove protection against virus, bacteria and fungi)</td> <td>X</td> <td>X</td> <td>X</td> </tr> </tbody> </table> <p>X = erforderlich / required</p>		5.1	5.2	5.3	Handschuh gegen Bakterien und Pilze (Glove protection against bacteria and fungi)	X	X		Handschuh gegen Viren, Bakterien und Pilze (Glove protection against virus, bacteria and fungi)	X	X	X		
	5.1	5.2	5.3												
Handschuh gegen Bakterien und Pilze (Glove protection against bacteria and fungi)	X	X													
Handschuh gegen Viren, Bakterien und Pilze (Glove protection against virus, bacteria and fungi)	X	X	X												
6	<b>Kennzeichnung</b> Marking Die Kennzeichnung von Schutzhandschuhen gegen gefährliche Chemikalien muss mit der Anforderung an Schutzhandschuhe in EN ISO 21420:2020 und mit folgenden Punkten übereinstimmen. All information shall be precise and comprehensive, and provided at least in the official language(s) of the country of destination.														
6.1 Bild / Fig. 2	<b>Kennzeichnung von Handschuhen des Typ A</b> Marking of Type A gloves Für Schutzhandschuhe, die die in 5.5 angegebenen Typ-A-Anforderungen erfüllen, ist das Piktogramm in Bild 2 mit Verweisung auf diesen Teil von ISO 374-1 zu verwenden. Die sechs getesteten Chemikalien müssen durch ihren Kennbuchstaben identifiziert werden, die unterhalb des Piktogramms angegeben werden müssen, wie in Bild 2 dargestellt. Wurden weitere Chemikalien geprüft, die nicht in der Liste angegeben sind, müssen die Informationen über die Leistungstufen in der Benutzeranleitung zur Verfügung gestellt werden. For protective gloves complying with the type A requirements stated in 5.5, the pictogram in Figure 2 shall be used with reference to this part of ISO 374-1. The six tested chemicals shall be identified by their code letter which shall be marked under the pictogram as shown in Figure 2. If other chemicals not present in the list have been tested, information about the performance levels shall be provided in the user instructions.	nicht anwendbar, da Typ C	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>												
6.2 Bild / Fig. 3	<b>Kennzeichnung von Handschuhen des Typ B</b> Marking of Type B gloves Für Schutzhandschuhe, die die in 5.5 angegebenen Typ-B-Anforderungen erfüllen, ist das Piktogramm in Bild 3 mit Verweisung auf diesen Teil von ISO 374-1 zu verwenden. Die drei getesteten Chemikalien müssen durch ihren Kennbuchstaben identifiziert werden, die unterhalb des	nicht anwendbar, da Typ C	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>												

THP\_EN ISO 374-1:2016 + A1:2018 (Rev.) V1.0




**Prüfbericht - Produkte**  
Test Report - Products

**TÜVRheinland**<sup>®</sup>

Prüfbericht-Nr.: Seite 10 von 28  
Test report no.: Page 10 of 28

Abzatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
5.4	Piktogramme angegeben worden müssen, wie in Bild 3 dargestellt. Wurden weitere Chemikalien geprüft, die nicht in der Liste angegeben sind, müssen die Informationen über die Leistungstufen in der Benutzeranleitung zur Verfügung gestellt werden. For protective gloves complying with the type B requirements stated in 5.5, the pictograms in Figure 3 shall be used with reference to the part of ISO 374-1. The three tested chemicals shall be identified by their code letter which shall be marked under the pictogram as shown in Figure 3. If other chemicals not present in the list have been tested, information about the performance levels shall be provided in the user instructions.		not applicable, because type C
6.3 Bild / Fig. 4	<b>Kennzeichnung von Handschuhen des Typ C</b> Marking of Type C gloves Für Schutzhandschuhe, die die in 5.5 angegebenen Typ-C-Anforderungen erfüllen, ist das Piktogramm in Bild 4 mit Verweisung auf diesen Teil von ISO 374-1 zu verwenden. Die getestete Chemikalie muss in der Gebrauchsanweisung mit Angaben zu ihrer Leistungsebene angegeben werden. Wurden weitere Chemikalien geprüft, die nicht in der Liste angegeben sind, müssen die Informationen über die Leistungstufen in der Benutzeranleitung zur Verfügung gestellt werden. For protective gloves complying with the type C requirements stated in 5.5, the pictogram in Figure 4 shall be used and the reference to this part of ISO 374-1. The tested chemical shall be given in the user instructions with information about its performance level. If other chemicals not present in the list have been tested, information about the performance levels shall be provided in the user instructions.	gegeben mit: BB74C BB74K1BB74C	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
EN 374-5	<b>Kennzeichnung Mikroorganismen</b> Marking microorganisms 6.2 Kennzeichnung von Handschuhen, die vor Bakterien und Pilzen schützen Marking of gloves protecting against bacteria and fungi Schutzhandschuhe die vor Mikroorganismen schützen, müssen den Anforderungen der EN ISO 21420:2020 Absatz 4, 5, und 7 entsprechen. Sie dürfen keine Beschädigung bei der Prüfung auf Wasser-Lack und Luft-Lack gemäß EN 374-2:2019 aufweisen. Protective gloves against micro-organism risks shall comply with the requirements given in EN ISO 21420:2020, Clause 4, Clause 5 and Clause 7.	gegeben, Handschuhe entsprechen Anforderungen der EN 21420:2020 Absatz 4.5 und 7 sowie weisen keine Beschädigung bei der Prüfung auf Wasser- und Luft-Lack auf	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>

THP\_EN ISO 374-1:2016 + A1:2018 (Rev.) V1.0



**Prüfbericht - Produkte**  
Test Report - Products

**TÜVRheinland**<sup>®</sup>

Prüfbericht-Nr.: Seite 11 von 28  
Test report no.: Page 11 of 28

Abzatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
	Protective gloves against virus, bacteria and fungi shall not leak when tested according to EN 374-2:2019, 7.2 Air-leak test and 7.3 weather-leak test. (ISO 374-2:2019)	damage during testing of water leak and air leak	
6.3	<b>Kennzeichnung von Handschuhen, die vor Viren, Bakterien und Pilzen schützen</b> Marking of gloves protecting against viruses, bacteria and fungi Schutzhandschuhe die vor Viren schützen, müssen den Anforderungen aus EN 374-5 6.2 entsprechen und dürfen gemäß ISO 18694 Verfahren B kein nachweisbare Transfer (c) (PPUM) des Phi-X174 Bakteriophage bei der Tier-Untersuchung aufweisen. Protective gloves against virus have to comply with the requirements of EN 374-5 6.2 and shall be tested according to ISO 18694 Procedure B and shall exhibit no detectable transfer (c) (PPUM) of the Phi-X174 bacteriophage in the assay (b). (ISO 374-5:2016)	nicht anwendbar, da nicht ausgeübt	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
	Protective gloves against virus have to comply with the requirements of EN 374-5 6.2 and shall be tested according to ISO 18694 Procedure B and shall exhibit no detectable transfer (c) (PPUM) of the Phi-X174 bacteriophage in the assay (b). (ISO 374-5:2016)	not applicable, because not marked	
7 EN 374-1	<b>Information des Herstellers</b> Information supplied by the manufacturer Die Informationen des Herstellers müssen in Übereinstimmung mit den Anforderungen an die Informationen stehen, die in EN ISO 21420 festgelegt sind.	gegeben	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>

THP\_EN ISO 374-1:2016 + A1:2018 (Rev.) V1.0



**Prüfbericht - Produkte**  
Test Report - Products

**TÜVRheinland**<sup>®</sup>

Prüfbericht-Nr.: Seite 12 von 28  
Test report no.: Page 12 of 28

Abzatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
	Sie müssen außerdem die Ergebnisse von 5.2, 5.3, 5.4 (Irritation, Degradation, Permeation) mitteilen, die Liste sämtlicher Chemikalien, auf die die Schutzhandschuhe geprüft wurden und die Leistungstufen, die bei der Permeationsprüfung erreicht wurden. Folgende Warnhinweise müssen in der Benutzeranleitung hinzugefügt werden: „Diese Information macht keine Angaben zur tatsächlichen Schutzdauer am Arbeitsplatz und zur Unterscheidung von Gemischen und reinen Chemikalien.“ „Der Widerstand gegen Chemikalien wurde unter Laborbedingungen an Proben beurteilt, die lediglich von der Handrinnenfläche entnommen wurden (ausgenommen ist der Fall, bei dem der Handschuh 400 mm oder länger ist). In diesem Fall wird oberhalb des Saubers geteilt und bezieht sich ausschließlich auf die getesteten Chemikalien. Er kann anders sein, wenn die Chemikalie in einem Gemisch verwendet wird.“ „Es wird eine Überprüfung empfohlen, ob die Handschuhe für die vorgesehene Verwendung geeignet sind, da die Bedingungen an Arbeitsplatz in Abhängigkeit von Temperatur, Abrieb und Degradation von denen der Typprüfung abweichen können.“ „Würden Schutzhandschuhe bereits verwendet, können sie aufgrund von Veränderungen ihrer physikalischen Eigenschaften geringeren Widerstand gegen gefährliche Chemikalien bieten. Durch bei Berührung mit Chemikalien verursachte Degradation, Bewegungen, Fadenziehen, Reibung usw. kann die tatsächliche Anwendungsdauer wesentlich reduziert werden. Bei aggressiven Chemikalien kann die Degradation der wichtigste Faktor sein, der bei der Auswahl von gegen Chemikalien beständigen Handschuhen zu berücksichtigen ist.“ „Vor der Anwendung sind die Handschuhe auf jegliche Fehler oder Mängel zu überprüfen.“ Bei Handschuhen, die mehrfach verwendet werden können, muss der Hersteller die relevanten Anleitungen für die Dekontamination angeben. Ist keine Information zur Dekontamination vorhanden, sind die Handschuhe nur für die einmalige Verwendung vorgesehen und folgende Warnhinweise ist hinzu zu fügen: „Nur für die einmalige Verwendung bestimmt.“ The information supplied by the manufacturer shall be in accordance with the requirements for information as defined in EN ISO 21420.	Das Schutzlevel ist mit Level 6 angegeben, (K) NaOH 40% Degradation bei -4,3% gegeben auf zusätzlichem Einleger gegeben in englisch gegeben in englisch gegeben in englisch gegeben in englisch gegeben in englisch n/a gegeben in englisch	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
	The information supplied by the manufacturer shall be in accordance with the requirements for information as defined in EN ISO 21420.	Die Schutzleistung ist gegeben mit Level 6, (K) NaOH 40%, Degradation at -4,3%	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>

THP\_EN ISO 374-1:2016 + A1:2018 (Rev.) V1.0












## INTERNATIONAL STANDARDS




### EN 374-1-2, EN 374-4-5, EN 21420

Prüfbericht - Produkte Test Report - Products		TÜVRheinland <sup>®</sup>	
Prüfbericht-Nr. : Test report no.:		Seite 21 von 28 Page 21 of 28	
Absatz Class	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
	<p>- If this property is claimed for a leather glove, it shall have a water vapour transmission of at least 0.2 mg/(cm<sup>2</sup>h) when tested according to 6.3.1.</p> <p>- If this property is claimed for a textile glove, it shall have a water vapour resistance less than or equal to 30 m<sup>2</sup>PaW when tested according to 6.3.2.</p>		
5.3.2	<p><b>Wasserdampfaufnahme</b> Water vapour absorption</p> <p>Wenn die schützenden Eigenschaften des Handschuhs die Wasserdampfdurchlässigkeit verhindern oder ausserordentlich, müssen Handschuhe, falls praktisch durchführbar, so konstant sein, dass die Schweldefaufnahme so stark wie möglich reduziert wird.</p> <p>Wenn diese Eigenschaft für einen Lederschuhs angegeben ist, muss er bei Prüfung nach 6.4.2 eine Wasserdampfaufnahmefähigkeit von <math>\geq 8 \text{ mg}/(\text{cm}^2 \text{h})</math> aufweisen.</p> <p>Where the protection characteristics of the glove inhibit or exclude water vapour transmission, when practicable, the glove shall be designed to reduce the permeation absorption as much as possible.</p> <p>If this property is claimed for a leather glove, it shall have a water vapour absorption of at least 8 mg/cm<sup>2</sup> for 8 h when tested according to 6.4.2.</p>	<p>nicht anwendbar, auf Grund der Schutzwirkung gegen gefährliche Chemikalien und Mikroorganismen</p> <p>not applicable, because of the protection against dangerous chemicals and micro-organisms</p>	<p>P <input type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input checked="" type="checkbox"/></p> <p>NT <input type="checkbox"/></p>
6	<b>Prüfverfahren</b> Test procedures		
7	<b>Kennzeichnung und Information</b> Marking and information		
7.1	<b>Allgemeines</b> General		
7.1.1	Alle Informationen müssen präzise und nachvollziehbar sein. All information shall be precise and comprehensive.	gegeben given	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>
7.2	<b>Kennzeichnung</b> Marking		
7.2.1	<b>Handschuhkennzeichnung</b> Glove marking		
7.2.1.1	Jeder Schutzhandschuh muss mit folgenden Angaben gekennzeichnet sein: a. Name, Handelsmarke oder andere Erkennungsmerkmale des Herstellers oder des bevollmächtigten Repräsentanten des Herstellers; b. Handschuhbezeichnung (Handelsname oder Code, die dem Benutzer die eindeutige Identifizierung des Produkts innerhalb des Sortiments des Herstellers bzw. des bevollmächtigten Repräsentanten ermöglichen);	Kennzeichnung gegeben auf der Verpackung Marking given on the packaging	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>







Prüfbericht - Produkte Test Report - Products		TÜVRheinland <sup>®</sup>	
Prüfbericht-Nr. : Test report no.:		Seite 22 von 28 Page 22 of 28	
Absatz Class	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
	<p>c. Größenbezeichnung;</p> <p>d. wenn der Handschuh einer oder mehreren spezifischen Normen entspricht (siehe Literaturnotizen), muss/müssen diese graphisch(n) Symbol(e) den Angaben in Anhang C entsprechen. Jedes graphische Symbol muss zusammen mit der Verweisung auf die anwendbare spezifische Norm und den Leistungsstufen angegeben werden (siehe 7.3.5), die stets in derselben feststehenden Reihenfolge angegeben werden müssen, die in der entsprechenden Norm festgelegt ist.</p> <p>e. Herstellungsdatum, zumindest Monat und Jahr (z. B. 2016/11), oder andere Mittel, mit denen die Rückverfolgbarkeit der Chargen sichergestellt wird;</p> <p>f. wenn anwendbar, das Ablaufdatum, zumindest Monat und Jahr (z. B. 2016/11), hinter dem graphischen Symbol der Sanduhr, wie in Anhang C dargestellt.</p> <p>Each protective glove shall be marked with the following information: a. Name, trade mark or other means of identification of the manufacturer or the manufacturer's authorized representative; b. Glove designation (commercial name or code allowing the user to identify clearly the product within the manufacturer's authorized representative's range); c. Size designation; d. Where the glove conforms to one or more specific standards (see Bibliography), the pictogram (s) shall be as specified in Annex C. Each pictogram shall be accompanied by the reference of the applicable specific standard and performance levels (see 7.3.5), which shall always be in the same fixed sequence as defined in the corresponding standard; e. Date of manufacturing, at least the month and year (for example 2016/11), or any mean ensuring the manufacturing batch traceability; g. If applicable, the obsolescence date, at least the month and year (for example 2016/11), behind the hour glass pictogram as shown in Annex C.</p>	<p>c. gegeben mit z.B. Gr. M (7,5)</p> <p>d. gegeben mit:  </p> <p>e. gegeben mit: </p> <p>f. gegeben mit: </p> <p>g. gegeben mit: </p>	
	<p>7.2.1.2 Die Kennzeichnung muss über die gesamte vorhersehbare Gebrauchsdauer deutlich sichtbar und lesbar angebracht sein. Kennzeichnungen oder Aufschriften, die zu Verwechslungen mit den obigen Kennzeichnungen führen könnten, dürfen nicht am Handschuh angebracht werden.</p> <p>The marking shall be affixed so as to be visible and legible throughout the foreseeable useful life of the glove. Marks or inscriptions which could be confused with the above marks shall not be affixed to the glove.</p>	<p>gegeben an der Verpackung, Kennzeichnung am Handschuh nicht möglich</p> <p>given on packaging, marking on glove not possible</p>	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>



Prüfbericht - Produkte Test Report - Products		TÜVRheinland <sup>®</sup>	
Prüfbericht-Nr. : Test report no.:		Seite 23 von 28 Page 23 of 28	
Absatz Class	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
7.2.1.3	<p>Solfern die Kennzeichnung auf dem Handschuh aufgrund der Produktgestaltung nicht möglich ist, ist sie auf der Verpackung oder einem dem Handschuh beiliegenden Dokument anzubringen.</p> <p>If marking on the glove is not possible given the characteristics of the product, the marking shall be affixed to the packaging or any document supplied with the glove.</p>	<p>gegeben, auf der Verpackung</p> <p>given, on packaging</p>	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>
7.2.1.4	<p>Ein graphisches Symbol darf nur angegeben werden, wenn der Handschuh die Mindestanforderungen der entsprechenden spezifischen Norm erfüllt.</p> <p>A pictogram shall only be used when the glove meets at least the minimum requirement of the relevant specific standard.</p>	<p>gegeben</p> <p>given</p>	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>
7.2.2	<b>Kennzeichnung der Verpackung</b> Marking of packaging		
	Jede Verpackung, die die Handschuhe unmittelbar enthält, muss eindeutig mit folgenden Angaben gekennzeichnet sein: a. Name und vollständige Anschrift des Herstellers oder des bevollmächtigten Repräsentanten des Herstellers; b. die in 7.2.1.1 b) und c) geforderten Informationen; c. Hinweis, wo die Informationen nach 7.3 erhalten werden können; d. wenn es sich um einen einfachen Handschuh handelt, der dem Benutzer nur gegen Gefahren wie die in Anhang A aufgeführten schützen soll, müssen die Worte „Nur für minimale Risiken“ oder eine äquivalente Formulierung aufgedruckt werden; e. das/die der spezifischen Norm entsprechende(n) graphisch(e) Symbol(e), siehe Anhang C, wenn der Handschuh dieser spezifischen Norm entspricht (siehe Literaturnotizen), jedem graphischen Symbol müssen eine Verweisung auf die entsprechende Norm sowie die Leistungsstufen hinzugefügt werden, und zwar immer in derselben Reihenfolge, wie sie in der zutreffenden spezifischen Norm festgelegt sind. Wenn zusätzliche graphische Symbole genutzt werden, müssen sie in den Informationen des Herstellers erläutert werden (7.3); f. sofern zutreffend, eine nach 7.3.6 geordnete Angabe; g. sofern anwendbar, das Ablaufdatum, zumindest Monat und Jahr (z. B. 2016/11), hinter dem graphischen Symbol der Sanduhr, wie in Anhang C dargestellt.	<p>a. gegeben</p> <p>b. Gr. M (7,5)</p> <p>c. gegeben mit Piktogramm</p> <p>d. n/a da PSA Kat. II</p> <p>e. gegeben mit:  </p> <p>f. n/a</p> <p>g. gegeben mit: </p>	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>



Prüfbericht - Produkte Test Report - Products		TÜVRheinland <sup>®</sup>	
Prüfbericht-Nr. : Test report no.:		Seite 24 von 28 Page 24 of 28	
Absatz Class	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
	<p>Each packaging enclosure that immediately contains the gloves shall be clearly marked with the following: A. Name and full address of the manufacturer or the manufacturer's authorized representative; B. The information required in 7.2.1.1 b) and c); C. References to where the information required in 7.3 may be obtained; d. Where the glove is of simple design intended to protect the wearer against only those hazards listed in Annex A, the words "For minimal risks only" or an equivalent expression shall be printed; e. The pictogram(s) appropriate to the specific standard, see Annex C, when the glove conforms to this specific standard (see Bibliography). Each pictogram shall be accompanied by the performance levels, which shall always be in the same fixed sequence as defined in the relevant specific standard, and the reference to the applicable standard. If additional pictograms are used, they shall be explained in the information supplied by the manufacturer (7.3); f. Where applicable, information required in 7.3.6; g. If applicable, the obsolescence date, at least the month and year (for example 2016/11), behind the hour glass pictogram as shown in Annex C.</p>	<p>a. given</p> <p>b. Gr. M (7,5)</p> <p>c. given by pictogram</p> <p>d. n/a - because PPE Cat. III</p> <p>e. given by:  </p> <p>f. n/a g. given by: </p>	
7.3	<b>Informationen des Herstellers</b> Information supplied by the manufacturer		
	<p>Wenn der Schutzhandschuh auf den Markt gebracht wird, müssen folgende Mindestinformationen bereitgestellt und verfügbar gehalten werden:</p> <p>The following minimum information shall be supplied when the protective glove is placed on the market and shall be maintained available.</p>	<p>gegeben</p> <p>given</p>	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>
7.3.1	Name und vollständige Anschrift des Herstellers oder seines bevollmächtigten Repräsentanten.	gegeben/given	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>
7.3.2	Handschuhbezeichnung nach 7.2.1.1 b) ; Glove designation as per 7.2.1.1 b).	gegeben/given	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>
7.3.3	Information zu dem verfügbaren Größenbereich und, sofern zutreffend, die nach 5.1 geforderten Informationen. Information on the available size range and where applicable, information required in 5.1.	gegeben mit S (8,5) bis XL (9) given by S (8,5) to XL (9)	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>



## INTERNATIONAL STANDARDS

### EN 374-1-2, EN 374-4-5, EN 21420

TÜVRheinland <sup>®</sup> Prüfbericht - Produkte Test Report - Products			
Prüfbericht-Nr.: Test report no.:		Seite 25 von 28 Page 25 of 28	
Abzatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
7.3.4	Der bestimmungsgemäße Gebrauch des Handschuhs und eine Verweisung auf die entsprechende(n) spezifische(n) Norm(en) und das Jahr der Veröffentlichung (siehe Literaturnummer).  The intended use of the glove and reference to the relevant specific standard(s) and publication year (see Bibliography).	gegeben in englisch  given by	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.5	Wenn nach 7.2.1.1 d) und 7.2.2 a) zutreffend, grafische Symbole, die die Gefahrenkategorien angeben, gegebenenfalls gefolgt von den Leistungsstufen.  0: gibt an, dass der Handschuh unter die Mindestleistungsstufe für eine bestimmte einzelne Gefahr fällt; X: gibt an, dass der Handschuh nicht geprüft wurde oder das Prüfverfahren für die Handschuhkonformitätsmessung oder das Handschuhmaterial ungeeignet scheint.  Weiterhin sind grundsätzliche Erklärungen beizufügen, um das Verstehen der relevanten Leistungsstufen zu unterstützen. Die Normen, auf die sie sich beziehen, sind anzugeben. Die Gründe für die Angabe „X“ müssen erklärt werden.  Die Leistungsstufen müssen in derselben Reihenfolge wie in der entsprechenden spezifischen Norm angegeben werden. Sie dürfen an einer beliebigen Stelle in der Nähe des grafischen Symbols angegeben werden, vorausgesetzt, sie stehen dazu in einem deutlichen Bezug.  Where applicable as per 7.2.1.1 d) and 7.2.2 a), pictograms indicating categories of hazard followed as applicable by the performance levels.  0: indicates that the glove falls below the minimum performance level for the given individual hazard. X: indicates that the glove has not been tested or the test method appears not to be suitable for the glove design or material.  Furthermore, a basic explanation shall be given to assist comprehension of the relevant performance levels, and the standards to which they refer shall be indicated. The reason(s) to use "X" shall be explained.  Performance level shall be in the same order as given within the relevant specific standard. They may be positioned anywhere near to the pictogram provided that they are in clear relation with it.	n/a: EN 374  Performance level  	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>



TÜVRheinland <sup>®</sup> Prüfbericht - Produkte Test Report - Products			
Prüfbericht-Nr.: Test report no.:		Seite 26 von 28 Page 26 of 28	
Abzatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
7.3.6	Wenn der Schutz nur auf einen Teil der Hand beschränkt ist, ist dies zu erwähnen.  When protection is limited to part of the hand only, this shall be mentioned.	n/a: da Schutz nicht nur auf einen Teil der Hand beschränkt ist  n/a: because protection is not only for a limited part of the glove	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.7	Sofern zutreffend, müssen Warnungen hinsichtlich möglicherweise eintretender Probleme oder Einschränkungen bei der Benutzung erwähnt werden. Beispielsweise könnte ein Warnhinweis zur Benutzung von rotfesten Handschuhen in der Nähe von drehenden Maschinenstellen gegeben werden.  If appropriate, warnings against problems likely to be encountered or limitation of use shall be mentioned. As an example, a warning could be given about the use of heat-resistant gloves used in close proximity of rotating machinery.	gegeben in englisch  given	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.8	Wenn die Materialien, aus denen der Handschuh besteht, die Leistungseigenschaften während der empfohlenen Lagerung beeinträchtigen könnten, müssen Informationen dazu angegeben werden, um sicherzustellen, dass durch die Lagerung die Eigenschaften des Handschuhs nicht wesentlich verändert werden.  If the materials constituting the gloves are known to lose their performance during recommended storage, information shall be given to ensure that the storage will not change the glove characteristics significantly.	gegeben in englisch  given by	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.9	Wenn die geplante Leistungsfähigkeit des Handschuhs durch die Alterung beeinträchtigt werden könnte, muss die erforderliche Angabe zur Festlegung eines angemessenen Abkühlungs- wie nach 7.2.1.1 f) gefordert, angegeben werden.  If it is known that the design performance of the glove may be significantly affected by aging, the necessary information to establish a reasonable obsolescence date as requested in 7.2.1.1 f) shall be given.	gegeben in englisch  given by	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.10	Bei Naturkautschuk enthaltenden Handschuhen ein Warnhinweis wie etwa „Der Handschuh enthält Naturkautschuk, der allergische Reaktionen hervorzurufen kann.“  A warning for gloves containing any natural rubber, such as: "The glove contains natural rubber which may cause allergic reactions."	n/a	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.11	Anweisungen zum Anziehen, Ausziehen und Richten der Handschuhe, Erhalten des Komforts und der Handhygiene, Schutz vor Kontamination der Hand und gegebenenfalls Angaben zur Kombination mit anderen PSA-Elementen.	gegeben in englisch	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>



TÜVRheinland <sup>®</sup> Prüfbericht - Produkte Test Report - Products			
Prüfbericht-Nr.: Test report no.:		Seite 27 von 28 Page 27 of 28	
Abzatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
7.3.12	Anweisungen relevant für Öffnen, Öffnen, Anpassen der Handschuhe, Erhalten des Komforts und Hygiene der Hand, Prävention vor Kontamination der Hand, und wenn relevant Information über Kombination mit anderen Formen von PPE.  Any relevant instruction to check the integrity of the glove before using it (for example check that the glove does not present holes, cracks, tears, colour change... and discard any glove presenting such defects).	gegeben  given	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.13	Anweisungen für die Lagerung.  Storage instructions.	gegeben in englisch  given by	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.14	Wenn die Reinigung nach ISO 3758 oder Erklärungen sowie eine annehmbare Anzahl an Reinigungszyklen angegeben werden.  Wenn keine Reinigung empfohlen wird, muss angegeben werden, dass der Handschuh nicht waschbar ist. Davon ausgenommen sind Einweghandschuhe.  If cleaning according to 4.3 is claimed, care symbols according to ISO 3758 or explanations and an acceptable number of cleaning cycles, shall be provided.  If cleaning is not recommended, it shall be indicated that the glove is not washable. This excludes single use gloves.	n/a: Einweghandschuh/ disposable gloves	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>
7.3.15	Gegebenenfalls Prüfergebnisse nach 4.4 zusammen mit einer Verweisung auf die entsprechende Norm, Prüfmethoden, Prüffläche des Handschuhs und das angewendete Prüfverfahren bzw. die genutzte Probestromschleife sowie die angelegte Prüfspannung nach der entsprechenden Norm. Darüber hinaus ist ein Warnhinweis anzugeben, dass die gesamte Bekleidung und alle Schuhe, die zusammen mit dieser Handschuh getragen werden, ebenfalls unter Berücksichtigung elektrostatischer Risiken geteilt sein müssen.  If relevant, test results according to 4.4 along with reference of corresponding standard, atmosphere for testing, area of the glove tested and test method/electrode used and the voltage applied as per the relevant standard. Moreover, a written warning shall be provided to advise that all clothing and shoes worn	n/a	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>



TÜVRheinland <sup>®</sup> Prüfbericht - Produkte Test Report - Products			
Prüfbericht-Nr.: Test report no.:		Seite 28 von 28 Page 28 of 28	
Abzatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
7.3.16	mit dieser Art von Handschuh sollte auch die elektrische Gefahr berücksichtigt werden.  Reference to accessories and spare parts, if relevant, for example connectors systems between hoses and gloves.	n/a	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>
7.3.17	Sofern relevant, die Art der für den Transport geeigneten Verpackung.  Type of packaging suitable for transport, if relevant.	n/a	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>
7.4	Auf Nachfrage bereitzustellende Information  Information to be supplied on request  Eine Liste der in dem Handschuh enthaltenen Stoffe, die bekanntermaßen Allergien verursachen, siehe Anhang G, muss auf Nachfrage bereitgestellt werden, mit Ausnahme von Naturkautschuk (7.3.10).  A list of the substances contained in the glove which are known to cause allergies, see Annex G, shall be supplied on request, other than natural rubber (7.3.10).	gegeben/given	informativ/ informative





## REGULATION COMPLIANCE

CE 0197



**CERTIFICATE**  
EU Type-Examination Certificate  
Regulation 2016/425/EU  
Personal Protective Equipment

Registration No.:  
Report No.:

Holder:

Product: Protective gloves against chemicals and micro-organisms according to EN ISO 374-1 + A1:2018 and EN ISO 374-5:2016

Identification: Disposable gloves  
Colour: blue  
Material: nitrile, wall thickness 0,05 mm  
Sizes: S (S.3) - XL (5)  
Performance parameters: Type C, class 6, K; NaOH 40%

- PPE Category III - obligatory monitoring module C2 -  
The EU type-examination certificate refers to the above mentioned product. This is to certify that the product complies with the essential requirements of Annex II of the regulation 2016/425/EU. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity. The holder is entitled to use this certificate in connection with the declaration of conformity in accordance with Annex IX.

Valid till: 07.03.2024

Date: 08.03.2019

  
 Notified Body  
 Dipl.-Ing. E. Albrecht

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg  
Notified by Zentralstelle der Länder für Sicherheitstechnik (ZLS).

Notified under No. 0197 to the EC Commission.

CE The CE marking may be used if all relevant and effective EC Directives are complied with. CE





Business Stream Products  
Textiles - PPE

Exactly Right.

TÜV Rheinland LGA Products GmbH - D-11109 Cologne Air Green Street

Your correspondence:  
Corinna Albrecht  
Tel. +49 201 800 5300  
Mail: tuvr@gm.lga.com  
curagrip@lga.com  
Cologne, 25 Jun 2022

Product surveillance acc. to Module C2 of PPE regulation (EU) 2016/425, annex VII

Your EU Type-Examination Certificate:  
Product: Protective gloves against dangerous chemicals and micro-organisms  
Article: Disposable gloves  
Certificate of conformity Type C  
Test requirements: EN ISO 374-1:2016+A1:2018 and EN ISO 374-5:2016 (EU) 2016/425  
Regulation:

Dear Mrs. Li,

The PPE Regulation (EU 2016/425) has fixed rules for the monitoring of manufactured PPE of category III with required in Annex VII (Module C2) "Conformity to type based on internal production control plus supervised product checks at random intervals".  
The submitted sample of the product has been tested and in this configuration found to be in accordance with the above mentioned requirements.

Enclosed please find the test report no.  
The certificate remain valid until 2024-03-07.  
Date of sample picking: 2022-03-31.

Best regards

X.C. Albrecht      x.H. Dr.

Certification: Signant Ion Corina Albrecht      Labmanager: Signant Ion Miriam Diefmeier

TÜV Rheinland LGA Products GmbH  
Am Gassen Stein  
51109 Cologne  
Germany  
Headquarter:  
TÜV Rheinland LGA Products GmbH  
Tillystraße 2  
90431 Nürnberg  
Germany  
Phone +49 911 800 5300  
Fax +49 911 800 5305  
www.tuv.com  
www.tuv.com/safety

Board of Management  
Dipl.-Ing.  
Jörg Müller, Spokesman  
Dipl.-Ing.  
Dr.-Ing. Seifmann  
Nürnberg HRB 200113  
VAT No. DE211920495  
Chairman of the  
Supervisory Board:  
Dr.-Ing. Michael Fink



## INTERNATIONAL STANDARDS

### EN 1186

Test Report No.: \_\_\_\_\_  
Report Date: 11 April 2019



**SUBJECT** Chemical Test

**TEST LOCATION** TÜV SÜD China  
TÜV SÜD Products Testing (Shanghai) Co., Ltd.  
B-34, No.1999 Du Hu Road, Minhang District  
Shanghai 201108, P.R. China

**CLIENT NAME**  
**CLIENT ADDRESS**

**TEST PERIOD** 01-Mar-2019-08-Mar-2019

**TEST REQUEST** In accordance with Council of Europe Res AP (2004) 4

**CONCLUSION** **PASS**  
The submitted sample was found to comply with the overall migration requirements as stated in European Resolution Res AP (2004) 4 on rubber to be used for food contact applications.

Prepared By: *Cynthia Cao*  
(Cynthia Cao)  
Report Drafter

Authorized By: *Le...*  
(Le...)  
Authorized Signatory





Note: (1) General Terms & Conditions as mentioned on order. (2) This report refers only to the items tested. (3) The test method is based on the standard as stated in full without the written approval of the laboratory. (4) Without the agreement of the laboratory, the client is not permitted to use the test results for any other purpose.

Chemical Technology Laboratory  
TÜV SÜD Products Testing (Shanghai) Co., Ltd.  
B-34, No.1999 Du Hu Road, Minhang District  
Shanghai 201108, P.R. China

Phone: +86 (21) 6011 6100  
Fax: +86 (21) 6011 6100  
Email: china@tuv.com.cn  
Website: www.tuv.com.cn

Regional Head Office:  
TÜV SÜD Certification and Testing  
Johann-Carl-Str. 140  
No. 21, 70564 Stuttgart  
70564 Stuttgart, P.R. China

TUV  
Page 1 of 2

Test Report No.: \_\_\_\_\_  
Report Date: 11 April 2019



**RECEIPT DATE / TEST DATE**  
01-Mar-2019 / 01-Mar-2019

**THE FOLLOWING SAMPLE(S) WAS/WERE SUBMITTED BY: ON BEHALF OF THE CLIENT(S) AS**

Sample Name: Powder free nitrile glove, blue  
Sample Specification: Medium  
Batch No./Order: /  
Manufacturer: /

SAMPLE NO.	DESCRIPTION	PHOTOGRAPH
	Blue glove	

**TEST METHOD(S)**

1. For material: Rubber  
- Overall migration test for compliance with European Resolution Res AP (2004) 4 on rubber to be used for food contact applications.  
- As specified in REGULATION (EU) No 10/2011 and its amendments, with reference to EN 1186: Part 2 (test methods for overall migration into olive oil by total immersion) / EN 1186: Part 3 (test methods for overall migration into aqueous food simulant by total immersion) / EN 1186: Part 14 (substitute test.)

**TEST RESULT(S)**

1. Overall Migration Test - with reference to EN 1186: Part 2, Part 3 & Part 14

Simulant(s) Used	Test Condition	Results [mg/kg]	Maximum Permissible Limit [mg/kg]
10% Ethanol	70 °C for 2 hours	<5.0	60
3% Acetic acid	70 °C for 2 hours	<5.0	60
90% Ethanol	60 °C for 2 hours	<5.0	60
Isobutylene	60 °C for 0.5 hour	<5.0	60

Note:  
1. might denote migration per kilogram foodstuff  
2. Specification is quoted from European Resolution Res AP (2004) 4 on rubber to be used for food contact applications  
3. < denotes less than

Note: This report is for internal use only by the client.

**END OF THE TEST REPORT**

Chemical Technology Laboratory  
TÜV SÜD Products Testing (Shanghai) Co., Ltd.  
B-34, No.1999 Du Hu Road, Minhang District  
Shanghai 201108, P.R. China

Phone: +86 (21) 6011 6100  
Fax: +86 (21) 6011 6100  
Email: china@tuv.com.cn  
Website: www.tuv.com.cn


Regional Head Office:  
TÜV SÜD Certification and Testing  
Johann-Carl-Str. 140  
No. 21, 70564 Stuttgart  
70564 Stuttgart, P.R. China

TUV  
Page 2 of 2




## REGULATION COMPLIANCE

### FDA 510K



**510(k) Premarket Notification**

510(k) | DRUGS | Registration & Listing | Adverse Events | Alerts | PUA | HCE | Classification | Statistics  
 CFR Title 21 | Radiation-Emitting Products | X-Ray Assembly | Vendor Reports | CLIA | TRC

**New Search** [Back To Search Results](#)

<b>Device Classification Name</b>	Polymer Patient Examination Glove
<b>510(k) Number</b>	K171873
<b>Device Name</b>	Powder Free Nitrile Patient Examination Glove, Blue Colored, Non Sterile, Tested For Use With Chemotherapy Drugs
<b>Applicant</b>	
<b>Applicant Contact Correspondent</b>	
<b>Correspondent Contact</b>	Da Shi
<b>Regulation Number</b>	
<b>Classification Product Code</b>	<a href="#">L2A</a>
<b>Subsequent Product Code</b>	<a href="#">L2C</a>
<b>Date Received</b>	06/23/2017
<b>Decision Date</b>	11/15/2017
<b>Decision</b>	Substantially Equivalent (SESE)
<b>Regulation Medical Speciality</b>	General Hospital
<b>510k Review Panel</b>	General Hospital
<b>Summary Type</b>	Summary Traditional
<b>Reviewed By Third Party</b>	No
<b>Combination Product</b>	No

Page Last Updated: 06/29/2021  
 Note: If you need help accessing information in different file formats, see Instructions for Downloading Viewers and Players.  
 Language Assistance Available: Español | 繁體中文 | Tiếng Việt | 한국어 | Tagalog | Πρωτοβουλία | العربية | Kreyòl Ayisyen | Français | Polski | Português | Italiano | Deutsch | 日本語 | العربية | English



## DECLARATION OF CONFORMITY

### Medical

---

### Declaration of Conformity

**Manufacturer:**  
**Address:**

**Product:** Disposable Nitrile Examination Gloves  
**Designation:** X-Small, Small, Medium, Large, X-Large, XX-Large


We herewith declare that the above-mentioned devices comply with the European Medical Device Regulation (EU) MDR 2017/745 and PPE Regulation (EU) 2016/425. The EU declaration of conformity is issued under the sole responsibility of the manufacturer.

By formulating the products, the chemical substances selecting was rigorous, and compliance to REACH, RoHS, Halogen-Free, SVHC IRL.

We strict following the standard of U.S. and EU, no DEHP, BBP, DBP, and DIBP is using in any vinyl products.

**STANDARDS**

Standards Harmonized Standards applicable to this product are:  
 EN455-1, EN455-2, EN455-3, EN455-4,  
 EN374-1, EN374-2, EN374-3, EN374-5.

**Signature:** 

**Date:** March 02, 2021



### Medical

### EU Declaration of Conformity

**Manufacturer:**  
**Address:**  
**SRN:**  
**European Representative:** Lotus NL B.V.  
 Konings Julianenplein 10, 1e Vloer, 2585AA The Hague, Netherlands  
**SRN:**  
**Product:** Disposable Medical nitrile exam glove  
 X-Small, Small, Medium, Large, X-Large  
**GMDN Code:**  
**UMDN Code:**  
**Basic UDI:**

**Classification (MDR, Annex VIII):** Class I, Rule 1.  
**Conformity Assessment Route:** EU DECLARATION OF CONFORMITY following the Annex II + Annex III + Article 19 of MDR (EU) 2017/745.

We herewith declare that the above mentioned de products meet the transposition into national law, the provisions of the following EU Regulation and Standards. All supporting documentations are retained under the premises of the manufacturer.  
 is exclusively responsible for the declaration of conformity.

**General applicable regulations, directives:**  
 Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC.  
**Applied standards, common specification, guidance:**  
 EN 455-1:2020, EN 455-2:2015, EN 455-3:2015, EN 455-4:2009

**Signature:** \_\_\_\_\_  
**Date:** July 06, 2021



### Chemical


### EU Declaration of Conformity - PPE

**The manufacturer:**

**Representative (EU, Switzerland):** CURADEN AG, Amlehnstrasse 22, 8010 Kriens, Switzerland

Declares under his sole responsibility, that the PPE reference described hereafter:

Protective nitrile Glove  
 PPE to be used against Category III risks  
 EN ISO 374-5:2016 EN ISO 374-1:2016/Type B



is in conformity with the provisions of Regulation (EU)2016/425 and with the European harmonized standards EN ISO 374-1:2016+A1:2018/Type B, EN ISO 374-5:2016, EN ISO 21420:2020 and is identical to the PPE which is subject to the EU-Type examination, under certificate number 2777, issued by the Notified Body:  
 SATRA Technology Europe Limited  
 Braostown Business Park, Clonee  
 D15YN2R, Republic of Ireland

and is subject to the procedure set out in Annex VII (Module C2) of the Regulation under the supervision of the Notified body:  
 SATRA Technology Europe Limited  
 Braostown Business Park, Clonee  
 D15YN2R, Republic of Ireland

**Signature:** \_\_\_\_\_

**Date:** January 04, 2022



### Food

### DECLARATION OF APPROVAL FOR THE USE OF PRODUCT WITH FOOD

**The Manufacturer:**

**Represented by (EU, Switzerland):** CURADEN AG, Amlehnstrasse 22, 8010 Kriens, Switzerland

**Declaration for the following model/gloves**

**Nitrile VELVET GRIP blue (Premium)**

**Relevant category:** "elastomers and rubber"

meets the following regulations:

**General principles of Europe Resolution (Res AP(2004) 4 on rubber, EC Regulations: 1055/2004 and 2020/2006 relating to good Manufacturing Practice for materials and articles intended to come into contact with food.**

All ingredients, starting monomers and additives used in the production of this glove are compliant with all positive lists - all relevant Specific Migration Limits (SML) or restrictions of the applicable EU food legislation.



## DECLARATION OF CONFORMITY

### Food

**Global Migration Data:**

Simulants used:	Test condition	Result(s) [mg/kg]	Max. allowed limit
Ethanol (WV) 16%	2 hours/ 70°C	<5.0	60
Acetic acid (WV) 3%	2 hours/ 70°C	<5.0	60
Ethanol (WV) 95%	2 hours/ 60°C	<5.0	60
iso-octane	0,5 hours/ 40°C	<5.0	60

According to EN 1185, the analytical tolerance for simulants of foods containing water, alcohol and acids is 1 mg/kg\*, for simulants of fatty foods 3 mg/kg\*.

**Storage instructions:** Protect from direct sunlight, store in a cool and dry place in the original packaging. Store away from sources of ozone. If the gloves are stored properly according to the above instructions, their performance and properties will not be significantly reduced. For gloves that may be affected by aging or storage, the use-by date is stated on the packaging materials.

Signature:

Date: March 16, 2022



PACKAGING

**VELVET GRIP**  
QUANTI IN NITRILE manopere

• Quantità di manopere  
• Non sterile  
• Ambidestro  
• Senza polvere e blu  
• Senza lattice

100 manopere

**VELVET GRIP**  
NITRILE Examination GLOVES

curaden  
better health for you

- Disposable Gloves
- Non-sterile
- Ambidextrous
- Powder-free and blue
- Latex-free

100 GLOVES

**VELVET GRIP**  
QUANTOS DE NITRILLO desechables

• Cantidad de manopere  
• No estériles  
• Ambidextros  
• Sin polvo y azul  
• Sin látex

100 guantes

**VELVET GRIP**  
Einweg NITRIL-HANDSCHÜHE

curaden  
better health for you

- Untersuchungshandschuh
- Nicht steril
- Beidseitig
- Pulverfrei und blau
- Latexfrei

100 HANDSCHÜHE

**VELVET GRIP**  
NITRILE Examination GLOVES

curaden  
better health for you

Importeur and Distributor  
**CURADEN AG**  
Königsplatz 22  
CH-6600 Olten, [www.curaden.ch](http://www.curaden.ch)

EN For medical use and treatment under hygienic conditions. Personal protective equipment. Nitrile disposable protective gloves. 3. Zweckbestimmung: für den ärztlichen Bereich und Behandlungen unter hygienischen Bedingungen. Persönliche Schutzausrüstung. Nitril Einweg Handschuhe. FR Usage prévu: équipement de protection individuelle. Latex de protection et stérile. Pour le secteur médical et pour l'usage des traitements dans les conditions d'hygiène. ES Uso previsto: equipo de protección personal. Guantes protectores desechables de nitrilo. Para el uso médico y para realizar procedimientos en condiciones higiénicas. IT Destinazione d'uso: dispositivi di protezione individuale. Guanti protettivi monouso in nitrile. Per l'uso sanitario medico e per trattamenti medici effettuati in condizioni igieniche. RU Назначение: средства индивидуальной защиты. Латекс-свободные одноразовые перчатки. Для применения в лечебных учреждениях и в условиях соблюдения санитарно-гигиенических требований. PL Przeznaczenie: środki ochrony osobistej. Jednorazowe rękawiczki nitrylowe. Do stosowania w branży medycznej w warunkach higienicznych. HR Namjena: sredstva zaštite osobne. Jednokratne rukavice od nitrilnog materijala. Prikladno za medicinsku praksu i liječenje u higijenskim uvjetima. BG Предназначение: средства за лична защита. Еднократно перчатки от нитрилен материал. Подходящи за медицинска практика и лечение в хигиенни условия. CZ Provozování: prostředky osobní ochrany. Použití: zdravotnické a léčebné účely. SK Použitie: prostriedky osobnej ochrany. Použitie: zdravotníckych a lekárske účely. PL Proszony: środki ochrony osobistej. Przeznaczenie: środki ochrony osobistej. Jednorazowe rękawiczki nitrylowe. Do stosowania w branży medycznej w warunkach higienicznych. PT Utilização: equipamentos de proteção individual. Para utilização em condições de higiene. Para utilização em áreas médicas e hospitalares. De utilização em condições de higiene. EN For medical use and treatment under hygienic conditions. Personal protective equipment. Nitrile disposable protective gloves. 3. Zweckbestimmung: für den ärztlichen Bereich und Behandlungen unter hygienischen Bedingungen. Persönliche Schutzausrüstung. Nitril Einweg Handschuhe. FR Usage prévu: équipement de protection individuelle. Latex de protection et stérile. Pour le secteur médical et pour l'usage des traitements dans les conditions d'hygiène. ES Uso previsto: equipo de protección personal. Guantes protectores desechables de nitrilo. Para el uso médico y para realizar procedimientos en condiciones higiénicas. IT Destinazione d'uso: dispositivi di protezione individuale. Guanti protettivi monouso in nitrile. Per l'uso sanitario medico e per trattamenti medici effettuati in condizioni igieniche. RU Назначение: средства индивидуальной защиты. Латекс-свободные одноразовые перчатки. Для применения в лечебных учреждениях и в условиях соблюдения санитарно-гигиенических требований. PL Przeznaczenie: środki ochrony osobistej. Jednorazowe rękawiczki nitrylowe. Do stosowania w branży medycznej w warunkach higienicznych. HR Namjena: sredstva zaštite osobne. Jednokratne rukavice od nitrilnog materijala. Prikladno za medicinsku praksu i liječenje u higijenskim uvjetima. BG Предназначение: средства за лична защита. Еднократно перчатки от нитрилен материал. Подходящи за медицинска практика и лечение в хигиенни условия. CZ Provozování: prostředky osobní ochrany. Použití: zdravotnické a léčebné účely. SK Použitie: prostriedky osobnej ochrany. Použitie: zdravotníckych a lekárske účely. PL Proszony: środki ochrony osobistej. Przeznaczenie: środki ochrony osobistej. Jednorazowe rękawiczki nitrylowe. Do stosowania w branży medycznej w warunkach higienicznych. PT Utilização: equipamentos de proteção individual. Para utilização em condições de higiene. Para utilização em áreas médicas e hospitalares. De utilização em condições de higiene.

4 260740 980185

(REF) ZHPFND2

CE Conformity with MDR 2017/2401 Class I

**LOTUS N. B.V.** Königsplatz 10, 1e Vert, 3585AA The Hague, Netherlands  
Tel: +31-70-4160990  
www.curaden.com/nl/en/gloves  
For more information on nitrile

ISO 13485:2016 ISO 9001:2015 EN ISO 11423:2020  
EN ISO 374-1:2014 A1:2018 Type B Level EN ISO 22746:2019  
9% Sodium hydroxide (K) 8 ± 0.3%  
32% Hydrogen peroxide (P) 2 24.1%  
33% Cornstarch (C) 4 34.3%

MADE IN PRC

**VELVET GRIP**  
QUANTI IN NITRILE manopere

• Quantità di manopere  
• Non sterile  
• Ambidestro  
• Senza polvere e blu  
• Senza lattice

100 manopere

CUR\_VG\_Z1\_CERT\_LN\_EN\_30012023\_v32

22 of 22